

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130125</u> ✓	
Name <u>GARY J. TELLEP</u>		ID# <u>2281</u>	Date <u>5-9-09</u>
A	Agency <u>AST</u>		Phone # <u>451-5100</u>
Instrument Location <u>1979 Peger Rd., FAIRBANKS</u> ✓			
Alco S/N <u>X124650</u>		Target Value <u>.080</u>	High Pressure <u>1150 psi</u>
B	Alco Test Value Average <u>.076</u> ✓		<u>.077</u> ✓
	1 st Alco	2 nd Alco	
Signature <u>Gary J. Tellep</u>		<u>CCB</u> <u>5/15/09</u>	
(OVER)			


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N

130125 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

MAY 09, 2009 ✓

OPERATOR'S NAME:

TELLEP GARY J

OPERATOR'S NUMBER: 2291

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

N

O.L. #: N

DEPT/AGENCY: FBK

CASE/REPORT: N

TEST TYPE: U

ALCO TARGET VALUE: .060 ✓

ALCO S/N: X124633

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.96 in		
ALCO TARGET	.000	00:16
BLANK TEST	.000	00:18
INTERNAL STANDARD	VERIFIED	00:18
ALCO TU 29.96 in	.076 ✓	00:18
BLANK TEST	.000	00:19
SUBJECT SAMPLE	.000	00:19
BLANK TEST	.000	00:20
ALCO TU 29.96 in	.077 ✓	00:20
BLANK TEST	.000	00:21

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

MAY 09, 2009 ✓

TIME 00:23

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	43c
BREATH TUBE:	42c
BAROMETER:	29.96 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abdefghijklmnop
qrstuvwxyz{|}~